



September 14, 2021

SCC Soft Computer
Attention: Mr. Kenton R. Smith
5400 Tech Data Drive
Clearwater, FL 33760

Re: BK210607

Trade/Device Name: SoftDonor Version 4.5.5.0
Regulation Number: 21 CFR 864.9165
Regulation Name: Blood establishment computer software and accessories
Regulatory Class: Class II
Product Code: MMH
Dated: June 7, 2021
Received: June 16, 2021

Dear Mr. Smith:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (the Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. Although this letter refers to your product as a device, please be aware that some cleared products may instead be combination products. The 510(k) Premarket Notification Database located at <https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpmn/pmn.cfm> identifies combination product submissions. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the **Federal Register**.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements

of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803) for devices or postmarketing safety reporting (21 CFR 4, Subpart B) for combination products (see <https://www.fda.gov/combination-products/guidance-regulatory-information/postmarketing-safety-reporting-combination-products>); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820) for devices or current good manufacturing practices (21 CFR 4, Subpart A) for combination products; and, if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <https://www.fda.gov/medical-devices/medical-device-safety/medical-device-reporting-mdr-how-report-medical-device-problems>.

For comprehensive regulatory information about medical devices and radiation-emitting products, including information about labeling regulations, please see Device Advice (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance>) and CDRH Learn (<https://www.fda.gov/training-and-continuing-education/cdrh-learn>). Additionally, you may contact the Division of Industry and Consumer Education (DICE) to ask a question about a specific regulatory topic. See the DICE website (<https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/contact-us-division-industry-and-consumer-education-dice>) for more information or contact DICE by email (DICE@fda.hhs.gov) or phone (1-800-638-2041 or 301-796-7100).

Sincerely,

Orieji Illoh -S

Digitally signed by Orieji Illoh -S
DN: c=US, o=U.S. Government, ou=HHS, ou=FDA, ou=People, cn=Orieji Illoh -S, 0.9.2342.19200300.100.1.1=2000420941
Date: 2021.09.14 16:25:19 -04'00'

Orieji Illoh, MD
Director
Division of Blood Components and Devices
Office of Blood Research and Review
Center for Biologics Evaluation and Research

Enclosure
Indications for Use

Indications for Use

510(k) Number: BK210607

Device Name: SoftDonor Version 4.5.5.0

Indications for Use:

The SoftDonor software application supports single and multi-site blood center facilities. The software application is intended to be used by knowledgeable, trained medical personnel to document, query, and view integrated information regarding donors and products. All steps and events are captured in this decision support system including the manufacture of blood and blood components, suitability/deferral of donors, and the release of products for transfusion purposes or for further manufacturing. The system supports the use of both user or donor administered history questionnaire functionality along with additional system logic to determine donor eligibility which includes testing as well as donor health screening. The system allows manual data entry and retrieval, as well as automated data exchange by interfacing to laboratory information systems, transfusion service systems, testing facilities and recruitment systems.

Prescription Use ✓ AND/OR Over-The-Counter Use _____
(Part 21 CFR 801 Subpart D) (21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CBER, Office of Blood Research and Review

Orieji Illoh -S Digitally signed by Orieji Illoh -S
DN: c=US, o=U.S. Government, ou=HHS,
ou=FDA, ou=People, cn=Orieji Illoh -S,
0.9.2342.19200300.100.1.1=2000420941
Date: 2021.09.14 16:37:11 -04'00'

Division Sign-Off
Office of Blood Research and Review